

Bruker 布鲁克顺磁共振用低温指状杜瓦

ER 167 FDS-Q Finger dewar



For particular, applications, like detection of nitroxide(NO), the nitrogen finger dewar, ER

167 FDS-Q, is the method of choice. The finger dewar minimizes sample preparation

time and experimental set-up. The sample temperature is at 77K.